



NABL

Department of Science & Technology, India

SCOPE OF ACCREDITATION

Laboratory	Belz Calibration Laboratory, 5L-123, N.I.T., Faridabad		
Accreditation Standard	ISO/IEC 17025:2005		
Field	Mechanical Calibration	Issue Date	01.07.2011
Certificate Number	C-0100	Valid Until	30.06.2013
Last Amended on	-	Page	9 of 9

Measured Qty / Instrument	Range	*Calibration Measurement Capability (\pm)	Remarks
6. <u>PNEUMATIC PRESSURE</u>			
ANALOGUE & DIGITAL PRESSURE GAUGES	0.5 bar to 10 bar	0.65% FS	Using Dig. Pressure Gauge
7. <u>HYDRUALIC PRESSURE</u>			
ANALOGUE & DIGITAL PRESSURE GAUGES	10 kg/cm ² to 700 kg/cm ²	0.25% FS	Using Digital Pressure Gauge

* Measurement Capability is expressed as an uncertainty (\pm) at a confidence probability of 95%

Convenor